

LEVEL-UN

Universal, Precision, Instrument Level

ISO 9000, GMP and other quality standards require instruments used for validation be calibrated and traceable to NIST standards. Most levels do not provide specific measured values, are not serialized and are not calibrated. Use of such levels does not provide useable data for validation purposes.



- *1.50 Inch Diameter easily fits the Dissolution Bath Base plate*
- *Includes serial number and one year calibration certification*
- *Bi-directional Bullet style does not need to be reoriented*
- *Sensitivity is .090 + or - .010 inch per foot (outer ring)*

Recommended pass fail criteria is: Dissolution instruments are adequately leveled when the bubble remains completely within the outer ring and does not touch the line.

Ordering Information:

Catalog Number	Description
LEVEL-UN	Universal, Precision, Instrument Level